



September 3, 2024

OSSTEM Implant Co., Ltd. Chair Business
% Peter Lee
Sr. Manager QA/RA Production
HIOSEN Inc.
85 Ben Fairless Dr.
Fairless Hills, Pennsylvania 19030

Re: K233805
Trade/Device Name: K5
Regulation Number: 21 CFR 872.6640
Regulation Name: Dental Operative Unit And Accessories
Regulatory Class: Class I, reserved
Product Code: EIA, KLC
Dated: August 14, 2024
Received: August 14, 2024

Dear Peter Lee:

We have reviewed your section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (the Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database available at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Additional information about changes that may require a new premarket notification are provided in the FDA guidance documents entitled "Deciding When to Submit a 510(k) for a Change to an Existing Device" (<https://www.fda.gov/media/99812/download>) and "Deciding When to Submit a 510(k) for a Software Change to an Existing Device" (<https://www.fda.gov/media/99785/download>).

Your device is also subject to, among other requirements, the Quality System (QS) regulation (21 CFR Part 820), which includes, but is not limited to, 21 CFR 820.30, Design controls; 21 CFR 820.90, Nonconforming product; and 21 CFR 820.100, Corrective and preventive action. Please note that regardless of whether a change requires premarket review, the QS regulation requires device manufacturers to review and approve changes to device design and production (21 CFR 820.30 and 21 CFR 820.70) and document changes and approvals in the device master record (21 CFR 820.181).

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR Part 803) for devices or postmarketing safety reporting (21 CFR Part 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR Part 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR Parts 1000-1050.

All medical devices, including Class I and unclassified devices and combination product device constituent parts are required to be in compliance with the final Unique Device Identification System rule ("UDI Rule"). The UDI Rule requires, among other things, that a device bear a unique device identifier (UDI) on its label and package (21 CFR 801.20(a)) unless an exception or alternative applies (21 CFR 801.20(b)) and that the dates on the device label be formatted in accordance with 21 CFR 801.18. The UDI Rule (21 CFR 830.300(a) and 830.320(b)) also requires that certain information be submitted to the Global Unique Device Identification Database (GUDID) (21 CFR Part 830 Subpart E). For additional information on these requirements, please see the UDI System webpage at <https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/unique-device-identification-system-udi-system>.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory->

[assistance/contact-us-division-industry-and-consumer-education-dice](#)) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Michael E. Adjodha -S

Michael E. Adjodha, MChE, RAC, CQIA
Assistant Director

DHT1B: Division of Dental and
ENT Devices

OHT1: Office of Ophthalmic, Anesthesia,
Respiratory, ENT, and Dental Devices

Office of Product Evaluation and Quality
Center for Devices and Radiological Health

Enclosure

Indications for Use

Submission Number (if known)

K233805

Device Name

K5

Indications for Use (Describe)

K5 is intended to supply power to and serve as a base for dental devices and accessories. This product includes a dental chair. The dental treatment unit is intended for use in the dental clinic environment and is used by trained dentists and/or dental assistants.

Type of Use (Select one or both, as applicable)



Prescription Use (Part 21 CFR 801 Subpart D)



Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."



Osstem Implant Co., Ltd. Chair Business

192, Haebong-ro, Danwon-gu, Ansan-si, Gyeonggi-do, 15428, Republic of Korea
Tel: +82 2 2016 7000 Fax: +82 2 2016 7001 www.osstem.com

510(k) Summary K233805

Date: September 2, 2024

1. Company and Correspondent making the submission

- Submitter's Name : Osstem Implant Co., Ltd. Chair Business
- Address : 192, Haebong-ro, Danwon-gu, Ansan-si, Gyeonggi-do, 15428, Republic of Korea
- Contact : Mr. Sangyong Lee
- Phone : +82-70-4626-0881

- Correspondent's Name : Hiossen Inc.
- Address : 85 Ben Fairless Dr. Fairless Hills, PA 19030
- Contact : Mr. Peter Lee
- Phone : +1-267-759-7031

2. Proposed Device

- Trade or (Proprietary) Name : K5
- Classification Name : Dental operative unit and accessories
- Regulation Number : 21 CFR872.6640
- Device Classification : Class I
- Classification Product Code : EIA

3. Predicated Device

- K3(K183347)

4. Description

- A Dental Unit and Chair K5 is designed for dental treatment of children and adults. It is a dental treatment unit in accordance with IEC 80601-2-60. The 3-way syringe is a dental instrument in accordance with EN 1639. It aids the dental application in the mouth of the patient by supplying air, water or spray. This product is designed for use in dentistry only and may only be used by trained medical personnel. K5 is similar to other commercially available products based on the intended use, the technology used, the claims, the electrical power and performance characteristics. It is substantially equivalent in design, function and intended use to the predicate device.

(1) Design Feature

K5, dental operative unit, has features of ergonomic and optimal position for treatment. It consists of adjusted touch panel, 3-way syringe, assistant touch panel, easy opening

and closing arm rest.

(2) Main Types

There are a couple of types for this device –K5 Mount and K5 Cart which consists of chair, unit, table, seat, stool, 3-way syringe, monitor arms, foot control and console.

(3) Technological Characteristics

K5 is an AC-powered dental operative unit with accessories, intended to supply power to and serve as a base for other dental devices. It includes a treatment chair, dentist element, assistant element and a dental light as offering several additional options and electronically-controlled chair movements with software and water unit functions.

5. Indication for use

K5 is intended to supply power to and serve as a base for dental devices and accessories. This product includes a dental chair. The dental treatment unit is intended for use in the dental clinic environment and is used by trained dentists and/or dental assistants.

6. Summary of Technological Characteristics

The K5 functions in a manner similar to and is intended for the same use to the predicate device. Primarily, K5 is substantially equivalent to the K3 (K183347) marketed by OSSTEM IMPLANT Co., Ltd and does have some different technological characteristics and slightly different external design. However, these differences do not raise new concerns of substantial equivalence as the performance data and testing of the K5 demonstrate that the devices are deemed to be substantially equivalent as described in a following comparison table:

Description	Subject Device	Predicate Device
Indications for Use	K5 is intended to supply power to and serve as a base for dental devices and accessories. This product includes a dental chair. The dental treatment unit is intended for use in the dental clinic environment and is used by trained dentists and/or dental assistants.	K3 is intended to supply power to and serve as a base for dental devices and accessories. This product includes a dental chair. The dental treatment unit is intended for use in the dental clinic environment and is used by trained dentists and/or dental assistants.
Product Name	K5	K3
Code	EIA	EIA
Power & Utility Supply	AC 100-120/220-240V, 50/60Hz, compressed air and water	AC 100-120/220-240V, 50/60Hz, compressed air and water
Main Components	Chair, Unit, Table, Seat, Stool,	Chair, Unit, Table, Seat, Stool,

	Monitor Arm, Hanaro Console	Monitor Arm*, Hanaro Console* (Note: K3 Cart* model applied ONLY)
Syringe	3-way syringe	3-way syringe
Control of water and air	Uses pneumatically controlled vales to water control the flow of air and water. On/off and intensity controlled by foot pedal.	Uses pneumatically controlled vales to water control the flow of air and water. On/off and intensity controlled by foot pedal.
Air Pressure	550kPa(min)/750kPa(max)	500kPa(min)/750kPa(max)
Water Pressure	250kPa(min)/600 kPa(max)	250kPa(min)/600 kPa(max)
Water System	City water supply	City water supply
Water Sanitation System	Distilled water container added	Distilled water container added
Suction	HVE (High volume evacuator) Saliva Ejectors	HVE (High volume evacuator) Saliva Ejectors
Patient Load	Max. 150kg	Max. 135kg
Chair Height	Max. 840±30mm, Min. 440±30mm	Max. 795±10mm, Min. 365±10mm
Back Rest	0°±5° to 70°±5°	0°±5° to 67°±5°
Head Rest	-90° to 70°	-10° to 45°
Lift Motor	Electromotor	Hydraulic electromotor
Electrical Safety	Complied with IEC 60601-1	Complied with IEC 60601-1
Electromagnetic compatibility	Complied with IEC 60601-1-2	Complied with IEC 60601-1-2

[Table 1] Comparison between K3 and its predicate device

7. Principle of Operation

Raising and lowering the dental chair

The K5 is an electronically controlled, electrically powered chair. The footswitch is used to position the chair and program auto-positioning functions into the chair. The chair hydraulic system is controlled by the electronics using electronic switching components and electrically powered solenoid valves.

Automatic positioning

Operating the 'Automatic' S/W on the Table Panel commands the chair's control device to automatically move and position the chair and seat to a predetermined position.

Hand-Piece Operation Circuit

Micro S/W, closing the supply air valve and opening the rubber film of the Master Block. Pressing the controller (Foot Control) Pedal at this time opens the supply air valve inside the Table, and the air of the compressor is supplied to the removed Hand-Piece through the

distribution block, thereby rotating the Hand-Piece Turbine. The air used to rotate the Hand-Piece Turbine is discharged to the atmosphere through the exhaust pipe.

Vacuum Operation

Removing the Suction from the Holder activates the Sensor mounted inside the Holder, opening the Air Solenoid Valve and supplying air to the suction device to enable suction.

3-Way Syringe

Pressing the button on the 3-way syringe, water or air is discharged.

Water Supply System

Upon detection of the cup by the Water Supply Port Sensor, PCB's Volume controls water volume.

Water Tube Cleaning

Pressing the Water Supply Button and Film Viewer Button together after supplying 1.2 L of cleaning solution (2~5% of hydrogen peroxide) to the Water Tube Cleaning/Distilled Water Container activates the Water Tube Cleaning System.

8. Non-clinical Test Data

Biocompatibility Testing

The biocompatibility evaluation for K5 components was conducted in accordance with the FDA Guidance Document and International Standard ISO 10993-1 "Biological Evaluation of Medical Devices –Part 1: Evaluation and Testing Within a Risk Management Process," as recognized by FDA. The waterlines and airlines were tested for cytotoxicity (ISO 10993-5). The components of the K5 are considered external communicating device for a duration of less than 24 hours.

Cleaning Validation

Cleaning/disinfection validation was conducted on the waterlines of the subject device. Validation was conducted using the following standards:

O 16954:2015 Dentistry – Test methods for dental unit waterline biofilm treatment

-Sampling for microbiological analysis

– Essential characteristics of test methods for the evaluation of treatment methods intended to improve or maintain the microbiological quality of dental unit procedural water. The microbiological simulation test and physical & chemical test are performed to ensure that the Dental chair waterline is effectively cleaned.

Reprocessing validation was conducted per FDA Guidance, "Reprocessing Medical Devices in Health Care Settings: Validation Methods and Labeling - Guidance for Industry and Food and Drug Administration Staff (fda.gov)" including cleaning, disinfection, and sterilization validation for the components of the K5 device.



Osstem Implant Co., Ltd. Chair Business

192, Haebong-ro, Danwon-gu, Ansan-si, Gyeonggi-do, 15428, Republic of Korea
Tel: +82 2 2016 7000 Fax: +82 2 2016 7001 www.osstem.com

Electrical Safety and Electromagnetic compatibility(EMC)

The Electrical Safety and Electromagnetic compatibility tests were performed in accordance with the following standards. Comprehensive performance testing has been conducted on the K5 in accordance FDA recognized standards. EMC testing was conducted in accordance with Standard IEC 60601-1-2. Electrical, mechanical, and environmental safety testing according to Standard IEC 60601-1 was performed. Usability testing was conducted in accordance with Standard IEC 60601-1-6 and IEC 62366.

Software and System Verification and Validation

Software verification and validation testing were conducted, and documentation was provided as recommended by FDA's Guidance for Industry and FDA Staff, "Guidance for the Content of Premarket Submissions for Software Contained in Medical Devices." The software for this device was considered as a "moderate" level of concern.

9. Clinical Test Data

No clinical studies are submitted.

No clinical studies are needed to characterize its performance and establish substantial equivalence.

10. Conclusion

Based upon the above mentioned data and comparison table, the K5 is substantially equivalent to the predicate device as described herein.