



December 6, 2024

Stryker Corporation (Tornier, Inc.)
Stefanie Tarara
Principal Regulatory Affairs Specialist
10801 Nesbitt Avenue South
Bloomington, Minnesota 55437

Re: K241878

Trade/Device Name: Tornier Humeral Reconstruction System (Tornier HRS); Tornier Perform
Humeral System - Stem (Tornier PHS-Stem)
Regulation Number: 21 CFR 888.3660
Regulation Name: Shoulder joint metal/polymer semi-constrained cemented prosthesis
Regulatory Class: Class II
Product Code: KWS, PHX, HSD
Dated: November 7, 2024
Received: November 8, 2024

Dear Stefanie Tarara:

We have reviewed your section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (the Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database available at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Additional information about changes that may require a new premarket notification are provided in the FDA guidance documents entitled "Deciding When to Submit a 510(k) for a Change to an Existing Device" (<https://www.fda.gov/media/99812/download>) and "Deciding When to Submit a 510(k) for a Software Change to an Existing Device" (<https://www.fda.gov/media/99785/download>).

Your device is also subject to, among other requirements, the Quality System (QS) regulation (21 CFR Part 820), which includes, but is not limited to, 21 CFR 820.30, Design controls; 21 CFR 820.90, Nonconforming product; and 21 CFR 820.100, Corrective and preventive action. Please note that regardless of whether a change requires premarket review, the QS regulation requires device manufacturers to review and approve changes to device design and production (21 CFR 820.30 and 21 CFR 820.70) and document changes and approvals in the device master record (21 CFR 820.181).

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR Part 803) for devices or postmarketing safety reporting (21 CFR Part 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR Part 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR Parts 1000-1050.

All medical devices, including Class I and unclassified devices and combination product device constituent parts are required to be in compliance with the final Unique Device Identification System rule ("UDI Rule"). The UDI Rule requires, among other things, that a device bear a unique device identifier (UDI) on its label and package (21 CFR 801.20(a)) unless an exception or alternative applies (21 CFR 801.20(b)) and that the dates on the device label be formatted in accordance with 21 CFR 801.18. The UDI Rule (21 CFR 830.300(a) and 830.320(b)) also requires that certain information be submitted to the Global Unique Device Identification Database (GUDID) (21 CFR Part 830 Subpart E). For additional information on these requirements, please see the UDI System webpage at <https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/unique-device-identification-system-udi-system>.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory->

[assistance/contact-us-division-industry-and-consumer-education-dice](#)) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Joseph P. Russell

Digitally signed by Joseph P.

Russell -S

Date: 2024.12.06 14:34:32 -05'00'

-S

for: Farzana Sharmin, PhD

Assistant Director

DHT6A: Division of Joint Arthroplasty Devices

OHT6: Office of Orthopedic Devices

Office of Product Evaluation and Quality

Center for Devices and Radiological Health

Enclosure

Indications for Use

Submission Number (if known)

K241878

Device Name

Tornier Humeral Reconstruction System (Tornier HRS);
Tornier Perform Humeral System - Stem (Tornier PHS-Stem)

Indications for Use (Describe)

Tornier Humeral Reconstruction System (Tornier HRS)

IN ANATOMIC:

The Tornier HRS is to be used only in patients with an intact or reconstructable rotator cuff, where it is intended to provide increased mobility and stability and to relieve pain. The Tornier HRS is indicated for use as a replacement of shoulder joints disabled by:

- Rheumatoid arthritis with pain
- Non-inflammatory degenerative joint disease (i.e. osteoarthritis and avascular necrosis)
- Correction of functional deformity
- Fractures of the humeral head
- Traumatic arthritis
- Revision of other devices if sufficient bone stock remains

IN REVERSE:

The Tornier HRS is indicated for use as a replacement of shoulder joints for patients with a functional deltoid muscle with pain disabled by:

- Rheumatoid arthritis
- Non-inflammatory degenerative joint disease (i.e. osteoarthritis and avascular necrosis)
- Correction of functional deformity
- Fractures of the humeral head
- Traumatic arthritis
- Massive and non-repairable rotator cuff tear
- Revision of other devices if sufficient bone stock remains

The reversed tray and polyethylene insert are indicated for use in the conversion from an anatomic to reversed shoulder arthroplasty without the removal of the humeral assembly during revision surgery for patients with a functional deltoid muscle.

Notes:

- All components are single use.
- The coated humeral stem is intended for cemented or cementless use.
- The all-poly glenoid components are intended for cemented use only
- The glenoid sphere implant is anchored to the bone with screws and is for non-cemented fixation.
- Titanium humeral heads are intended for patients with suspected cobalt alloy material sensitivity. The wear properties of Titanium and Titanium alloys are inferior to that of cobalt alloy. A Titanium humeral head is not recommended for patients who lack a suspected material sensitivity to cobalt alloy.

Tornier Perform Humeral System - Stem (Tornier PHS-Stem)

In Anatomic:

The humeral stem, humeral head coupler and humeral head may be used by themselves, as a hemiarthroplasty, if the natural glenoid provides a sufficient bearing surface, or in conjunction with a glenoid implant, as a total replacement.

The PERFORM™ Humeral System – Stem is to be used only in patients with an intact or reconstructable rotator cuff, where it is intended to provide increased mobility, stability, and to relieve pain. The PERFORM Humeral System – Stem is indicated for use as a replacement of shoulder joints disabled by:

- Non-inflammatory degenerative joint disease (i.e. osteoarthritis) and avascular necrosis
- Proximal humeral malunions
- Post-traumatic arthritis
- Revisions or fractures of the humeral head where adequate fixation can be achieved and adequate bone stock remains

Titanium humeral heads are intended for patients with suspected cobalt alloy material sensitivity. The wear properties of titanium and titanium alloys are inferior to that of cobalt alloy. A titanium humeral head is recommended for patients with a suspected material sensitivity to cobalt alloy.

All components are single use. The humeral stems are intended for cemented or cementless use.

The PERFORM Humeral System – Stem is intended to be used with cemented polyethylene glenoid components, in a total shoulder arthroplasty.

In Reverse:

The PERFORM™ Humeral System – Stem is indicated for use as a replacement of a shoulder joint for patients with a functional deltoid muscle, grossly deficient rotator cuff, and pain disabled by one or more of the following:

- Non-inflammatory degenerative joint disease (i.e. osteoarthritis) and avascular necrosis
- Pseudoparalysis or anterior superior escape
- Rotator cuff tear arthropathy
- Proximal humeral malunions
- Post-traumatic arthritis
- Revisions or fractures of the humeral head where adequate fixation can be achieved and adequate bone stock remains

The reversed insert is indicated for use for the conversion from an anatomic to reverse shoulder prosthesis without the removal of a well fixed humeral stem for patients with a functional deltoid muscle.

All components are single use. The humeral stems are intended for cemented or cementless use. The PERFORM Humeral System – Stem is intended to be used with glenoid implants that are anchored to the bone with screws for non-cemented fixation.

Type of Use (Select one or both, as applicable)



Prescription Use (Part 21 CFR 801 Subpart D)



Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

Contact Details

[21 CFR 807.92\(a\)\(1\)](#)

Applicant Name	Stryker Corporation (Tornier, Inc.)
Applicant Address	10801 Nesbitt Avenue South Bloomington MN 55437 United States
Applicant Contact Telephone	1-269-800-2754
Applicant Contact	Mrs. Stefanie Tarara
Applicant Contact Email	stefanie.tarara@stryker.com
Correspondent Name	Stryker Corporation (Tornier, Inc.)
Correspondent Address	10801 Nesbitt Avenue South Bloomington MN 55437 United States
Correspondent Contact Telephone	1-269-800-2754
Correspondent Contact	Mrs. Stefanie Tarara
Correspondent Contact Email	stefanie.tarara@stryker.com

Device Name

[21 CFR 807.92\(a\)\(2\)](#)

Device Trade Name	Tornier Humeral Reconstruction System (Tornier HRS); Tornier Perform Humeral System - Stem (Tornier PHS-Stem)
Common Name	Prosthesis, Shoulder, Semi-Constrained, Metal/Polymer Cemented
Classification Name	Shoulder joint metal/polymer semi-constrained cemented prosthesis
Regulation Number	888.3660
Product Code(s)	KWS, PHX, HSD

Legally Marketed Predicate Devices

[21 CFR 807.92\(a\)\(3\)](#)

Predicate #	Predicate Trade Name (Primary Predicate is listed first)	Product Code
K191318	Aequalis Flex Revive Shoulder System	KWS
K201315	Perform Humeral System - Stem (PHS-Stem)	KWS

Device Description Summary

[21 CFR 807.92\(a\)\(4\)](#)

Tornier HRS (formerly branded Aequalis Flex Revive Shoulder System) is a fully convertible anatomic and reversed shoulder arthroplasty system that is designed to be used with existing Tornier implant systems. It is a non-constrained system intended for total or partial replacement of the glenohumeral articulation.

The Tornier Humeral Reconstruction System (Tornier HRS) is a line extension of the existing Aequalis™ Flex Revive™ Shoulder System (AFR) (K191318, cleared June 14, 2019) that builds upon the AFR system with additional sized distal stems, proximal bodies, and reversed insert trays, new monoblock stems, humeral head couplers and MR Conditional labeling for Tornier HRS line extension components.

The Tornier Perform Humeral System-Stem (Tornier PHS-Stem) (K201315) is also being expanded to include additional sized Vitamin E UHMWPE reversed inserts and a renaming of these inserts for clarity.

Intended Use/Indications for Use

[21 CFR 807.92\(a\)\(5\)](#)

Tornier Humeral Reconstruction System (Tornier HRS)

IN ANATOMIC:

The Tornier HRS is to be used only in patients with an intact or reconstructable rotator cuff, where it is intended to provide increased mobility and stability and to relieve pain. The Tornier HRS is indicated for use as a replacement of shoulder joints disabled by:

- Rheumatoid arthritis with pain
- Non-inflammatory degenerative joint disease (i.e. osteoarthritis and avascular necrosis)
- Correction of functional deformity
- Fractures of the humeral head
- Traumatic arthritis
- Revision of other devices if sufficient bone stock remains

IN REVERSE:

The Tornier HRS is indicated for use as a replacement of shoulder joints for patients with a functional deltoid muscle with pain disabled by:

- Rheumatoid arthritis
- Non-inflammatory degenerative joint disease (i.e. osteoarthritis and avascular necrosis)
- Correction of functional deformity
- Fractures of the humeral head
- Traumatic arthritis
- Massive and non-repairable rotator cuff tear
- Revision of other devices if sufficient bone stock remains

The reversed tray and polyethylene insert are indicated for use in the conversion from an anatomic to reversed shoulder arthroplasty without the removal of the humeral assembly during revision surgery for patients with a functional deltoid muscle.

Notes:

- All components are single use.
- The coated humeral stem is intended for cemented or cementless use.
- The all-poly glenoid components are intended for cemented use only
- The glenoid sphere implant is anchored to the bone with screws and is for non-cemented fixation.
- Titanium humeral heads are intended for patients with suspected cobalt alloy material sensitivity. The wear properties of Titanium and Titanium alloys are inferior to that of cobalt alloy. A Titanium humeral head is not recommended for patients who lack a suspected material sensitivity to cobalt alloy.

Tornier Perform Humeral System - Stem (Tornier PHS-Stem)

In Anatomic:

The humeral stem, humeral head coupler and humeral head may be used by themselves, as a hemiarthroplasty, if the natural glenoid provides a sufficient bearing surface, or in conjunction with a glenoid implant, as a total replacement.

The PERFORM™ Humeral System – Stem is to be used only in patients with an intact or reconstructable rotator cuff, where it is intended to provide increased mobility, stability, and to relieve pain. The PERFORM Humeral System – Stem is indicated for use as a replacement of shoulder joints disabled by:

- Non-inflammatory degenerative joint disease (i.e. osteoarthritis) and avascular necrosis
- Proximal humeral malunions
- Post-traumatic arthritis
- Revisions or fractures of the humeral head where adequate fixation can be achieved and adequate bone stock remains

Titanium humeral heads are intended for patients with suspected cobalt alloy material sensitivity. The wear properties of titanium and titanium alloys are inferior to that of cobalt alloy. A titanium humeral head is recommended for patients with a suspected material sensitivity to cobalt alloy.

All components are single use. The humeral stems are intended for cemented or cementless use.

The PERFORM Humeral System – Stem is intended to be used with cemented polyethylene glenoid components, in a total shoulder arthroplasty.

In Reverse:

The PERFORM™ Humeral System – Stem is indicated for use as a replacement of a shoulder joint for patients with a functional deltoid muscle, grossly deficient rotator cuff, and pain disabled by one or more of the following:

- Non-inflammatory degenerative joint disease (i.e. osteoarthritis) and avascular necrosis
- Pseudoparalysis or anterior superior escape
- Rotator cuff tear arthropathy
- Proximal humeral malunions
- Post-traumatic arthritis
- Revisions or fractures of the humeral head where adequate fixation can be achieved and adequate bone stock remains

The reversed insert is indicated for use for the conversion from an anatomic to reverse shoulder prosthesis without the removal of a well fixed humeral stem for patients with a functional deltoid muscle.

All components are single use. The humeral stems are intended for cemented or cementless use. The PERFORM Humeral System – Stem is intended to be used with glenoid implants that are anchored to the bone with screws for non-cemented fixation.

Indications for Use Comparison

[21 CFR 807.92\(a\)\(5\)](#)

The subject device does not raise different questions of safety or effectiveness since the indications for use are identical to the predicate devices.

Technological Comparison

[21 CFR 807.92\(a\)\(6\)](#)

The Tornier HRS and the predicate Aequalis Flex Revive Shoulder System have the same intended use, same principle of operation, and similar technological features. Differences for the subject Tornier HRS include: Even sized distal stem diameters, smaller monoblock stems, proximal body with 135 degree neck angle, humeral head couplers in three offset options, reversed tray sizes, and the addition of MR Conditional labeling.

Non-Clinical and/or Clinical Tests Summary & Conclusions

[21 CFR 807.92\(b\)](#)

Non-clinical testing was performed to demonstrate substantial equivalence to the predicate device.

- Fatigue testing with corrosion evaluation
- Locking mechanism fatigue testing
- Pull-out and torque-out testing
- Wear and range of motion testing
- Material characterization testing
- Biocompatibility evaluation
- Packaging and shelf-life evaluations
- Distribution testing
- Sterilization evaluation
- Endotoxin testing
- MRI compatibility evaluation
- System Compatibility Testing

No clinical studies were performed.

The Tornier HRS does not raise different questions of safety or effectiveness. Differences in technological characteristics have been addressed with performance testing. The results of performance testing for the Tornier HRS support substantial equivalence to the predicate Aequalis Flex Revive Shoulder System (K191318, cleared on June 14, 2019).