



February 28, 2025

Promethean Restorative LLC
Glenn Bowman
President and CEO
333 Perry Street, Suite 210
Castle Rock, Colorado 80104

Re: K243565

Trade/Device Name: DYNAMIS™ SI Screw System
Regulation Number: 21 CFR 888.3040
Regulation Name: Smooth Or Threaded Metallic Bone Fixation Fastener
Regulatory Class: Class II
Product Code: OUR, HWC
Dated: January 29, 2025
Received: January 29, 2025

Dear Glenn Bowman:

We have reviewed your section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (the Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database available at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Additional information about changes that may require a new premarket notification are provided in the FDA guidance documents entitled "Deciding When to Submit a 510(k) for a Change to an Existing Device" (<https://www.fda.gov/media/99812/download>) and "Deciding When to Submit a 510(k) for a Software Change to an Existing Device" (<https://www.fda.gov/media/99785/download>).

Your device is also subject to, among other requirements, the Quality System (QS) regulation (21 CFR Part 820), which includes, but is not limited to, 21 CFR 820.30, Design controls; 21 CFR 820.90, Nonconforming product; and 21 CFR 820.100, Corrective and preventive action. Please note that regardless of whether a change requires premarket review, the QS regulation requires device manufacturers to review and approve changes to device design and production (21 CFR 820.30 and 21 CFR 820.70) and document changes and approvals in the device master record (21 CFR 820.181).

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR Part 803) for devices or postmarketing safety reporting (21 CFR Part 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR Part 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR Parts 1000-1050.

All medical devices, including Class I and unclassified devices and combination product device constituent parts are required to be in compliance with the final Unique Device Identification System rule ("UDI Rule"). The UDI Rule requires, among other things, that a device bear a unique device identifier (UDI) on its label and package (21 CFR 801.20(a)) unless an exception or alternative applies (21 CFR 801.20(b)) and that the dates on the device label be formatted in accordance with 21 CFR 801.18. The UDI Rule (21 CFR 830.300(a) and 830.320(b)) also requires that certain information be submitted to the Global Unique Device Identification Database (GUDID) (21 CFR Part 830 Subpart E). For additional information on these requirements, please see the UDI System webpage at <https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/unique-device-identification-system-udi-system>.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

**Maziar Shah
Mohammadi**

Digitally signed by Maziar Shah
Mohammadi

Date: 2025.02.28 10:26:28
-05'00'

For: Colin O'Neill, M.B.E.

Assistant Director

DHT6B: Division of Spinal Devices

OHT6: Office of Orthopedic Devices

Office of Product Evaluation and Quality

Center for Devices and Radiological Health

Enclosure

Indications for Use

Submission Number (if known)

K243565

Device Name

DYNAMIS™ SI Screw System

Indications for Use (Describe)

The DYNAMIS™ SI Screw System is indicated for sacroiliac joint fusion for:

- Sacroiliac joint dysfunction including sacroiliac joint disruption and degenerative sacroiliitis.
- Augmenting immobilization and stabilization of the sacroiliac joint in skeletally mature patients undergoing sacropelvic fixation as part of a lumbar or thoracolumbar fusion.

The DYNAMIS™ SI Screw System is also indicated for fracture fixation of the pelvis, including acute, non-acute and nontraumatic fractures.

Type of Use (Select one or both, as applicable)



Prescription Use (Part 21 CFR 801 Subpart D)



Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary

In accordance with Title 21 of the Code of Federal Regulations, Part 807, and in particular 21 CFR §807.92, the following summary of information is provided:

Company: Promethean Restorative LLC
333 Perry Street, Suite 210
Castle Rock, CO 80104

Contact: Glenn Bowman
Promethean Restorative LLC
333 Perry Street, Suite 210
Castle Rock, CO 80104
Phone: 303-330-8773
regulatory@prometheanrestorative.com

Date Prepared: February 26, 2025

Device Trade Name: DYNAMIS™ SI Screw System
Common Name: Sacroiliac Joint Fixation
Classification: 21 CFR §888.3040
Classification Name: Smooth or threaded metallic bone fixation fastener
Class: II
Product Code: OUR, HWC

Primary Predicate: SI-BONE iFuse TORQ® Implant System (K241574)
Additional Predicates: Curiteva, Sacroiliac Joint Fusion System (K210402)
Genesys Spine Sacroiliac Joint Fusion System (K191748)

Device Description:

The DYNAMIS™ SI Screw System consists of threaded, fenestrated, cannulated, 3D-printed implants and associated instruments. Implants are constructed from medical grade titanium alloy (Ti-6Al-4V ELI per ASTM F3001). The implants are fully threaded with a lag design. To accommodate varying patient anatomy, the DYNAMIS™ SI Screw Implants are available in multiple diameters and length offerings. The DYNAMIS™ SI Screw Implants consist of a cannulated central threaded body. The implants allow for packing of autograft and allograft materials. Using the designated instrument system, two or more implants should be inserted across the SI Joint to apply a compressive force across the joint and to provide stabilization and fusion. The DYNAMIS™ SI Screw implants are single use devices that are provided sterile.

Indications For Use:

The DYNAMIS™ SI Screw System is indicated for sacroiliac joint fusion for:

- Sacroiliac joint dysfunction including sacroiliac joint disruption and degenerative sacroiliitis.
- Augmenting immobilization and stabilization of the sacroiliac joint in skeletally mature patients undergoing sacropelvic fixation as part of a lumbar or thoracolumbar fusion.

The DYNAMIS™ SI Screw System is also indicated for fracture fixation of the pelvis, including acute, non-acute and nontraumatic fractures.



Substantial Equivalence:

The DYNAMIS™ SI Screw System has been demonstrated to be substantially equivalent with respect to indications, design, materials, function, manufacturing, and performance as compared to the predicate devices.

Performance Data:

Testing on the DYNAMIS™ SI Screw System included static cantilever bending and dynamic cantilever bending per ASTM F3574, as well as static axial pullout, static torsion, and driving torque per ASTM F543. The results demonstrate that the DYNAMIS™ SI Screw System is substantially equivalent to the predicate devices.

Conclusion:

Based on the indications for use, technological characteristics, performance testing, and comparison to predicate devices, the subject DYNAMIS™ SI Screw System has been shown to be substantially equivalent to legally marketed predicate devices.